
Midmark M11 - v1.0.0
Total Cycles: 1501

Name: RITTER 1
Sterilizer ID: V2178198
01/06/2022 10:55 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.5 F 0.0
2:00 97.8 F 0.0
4:00 116.4 F 0.1
6:00 146.7 F 1.8
8:00 174.4 F 4.8
10:00 195.9 F 9.5
12:00 213.8 F 6.9
14:00 227.3 F 7.8
16:00 239.4 F 12.7
18:00 238.1 F 9.8
20:00 236.6 F 8.3
22:00 249.4 F 14.7
24:00 261.5 F 21.6
26:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.5 F 27.9
2:00 271.9 F 28.0
3:00 272.4 F 28.4

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 231.5 F 0.0
5:00 211.4 F 0.0
Dry Cycle Aborted

FILLING: 0:60
HEATING: 26:19
STERILIZING: 3:00
VENTING: 2:48
DRYING: 8:57
TOTAL CYCLE: 00:42:03

01/06/2022 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____