
Midmark M11 - v1.0.0
Total Cycles: 1974

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2022 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 150.3 F 2.9
4:00 166.7 F 6.5
6:00 193.3 F 11.1
8:00 212.9 F 4.8
10:00 225.5 F 10.0
12:00 232.8 F 7.9
14:00 239.4 F 11.6
16:00 237.6 F 9.4
18:00 251.0 F 16.0
20:00 262.5 F 22.6
22:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.8 F 28.6
3:00 271.9 F 29.1
4:00 272.4 F 29.1

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 214.9 F 0.0
10:00 236.9 F 0.0
15:00 231.9 F 0.0
20:00 224.1 F 0.0
25:00 216.4 F 0.0
30:00 234.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

05/30/2022 03:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____