
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004885
14:30 21/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:26
STERILIZING 3:24
136.4 C 221kPa 3:24
136.2 C 219kPa 3:54
136.4 C 223kPa 4:24
136.0 C 220kPa 4:54
136.2 C 221kPa 5:24
136.1 C 220kPa 5:54
136.3 C 221kPa 6:24
136.2 C 221kPa 6:54
136.2 C 221kPa 6:54
VENTING 6:54
AIR DRYING 7:44
CYCLE COMPLETE 12:00

Digital Signature #
3027B38DEBFA1DB0
