
Midmark M11 - v1.0.0
Total Cycles: 2191

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.2 F 0.0
2:00 152.0 F 0.2
4:00 155.7 F 3.6
6:00 184.9 F 7.6
8:00 208.4 F 12.3
10:00 218.2 F 5.7
12:00 230.6 F 10.3
14:00 233.7 F 7.9
16:00 240.8 F 11.9
18:00 237.4 F 9.6
20:00 250.6 F 15.8
22:00 261.7 F 22.1
24:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.3
2:00 271.8 F 28.6
3:00 271.7 F 29.1
4:00 272.4 F 29.2

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 225.7 F 0.0
10:00 227.6 F 0.0
15:00 227.6 F 0.0
20:00 230.0 F 0.0
25:00 227.3 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:33
CYCLE COMPLETE

08/03/2022 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____