
Midmark M11 - v1.0.5
Total Cycles: 2483

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 07:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 75.4 F 0.0
4:00 82.4 F 0.0
6:00 113.6 F 0.0
8:00 150.8 F 1.9
10:00 175.0 F 4.8
12:00 192.9 F 8.4
14:00 213.1 F 12.4
16:00 226.5 F 6.5
18:00 236.1 F 10.9
20:00 230.9 F 7.6
22:00 243.0 F 12.7
24:00 240.6 F 11.0
26:00 253.1 F 17.3
28:00 263.8 F 23.7
30:00 271.0 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.2 F 29.2
2:00 272.7 F 29.6
3:00 272.5 F 29.9
4:00 271.4 F 29.4

Min 271.3 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 229.8 F 0.0
10:00 225.2 F 0.0
15:00 178.1 F 0.0
20:00 175.4 F 0.0
25:00 174.9 F 0.0
30:00 174.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

06/08/2023 08:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____