

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002006  
13:37 10/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:06  
STERILIZING 3:06  
136.0 C 221kPa 3:06  
136.4 C 226kPa 3:36  
136.3 C 225kPa 4:06  
136.4 C 223kPa 4:36  
136.4 C 222kPa 5:06  
136.3 C 225kPa 5:36  
136.4 C 223kPa 6:06  
136.4 C 225kPa 6:36  
136.4 C 225kPa 6:36  
VENTING 6:37  
AIR DRYING 7:33  
CYCLE COMPLETE 37:33

-----  
Digital Signature #  
02328ABD456EBC59  
-----