
Midmark M11 - v1.0.0
Total Cycles: 1131

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2021 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.4 F 0.0
2:00 105.6 F 0.0
4:00 135.4 F 0.1
6:00 166.9 F 1.3
8:00 189.3 F 5.2
10:00 215.0 F 9.3
12:00 228.5 F 12.8
14:00 230.2 F 8.1
16:00 239.3 F 12.9
18:00 238.0 F 10.2
20:00 235.9 F 8.4
22:00 249.4 F 14.9
24:00 261.2 F 21.5
26:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.3 F 28.8
2:00 272.7 F 28.7
3:00 271.6 F 28.2
4:00 273.0 F 29.2

Min 271.6 F 28.1
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 237.3 F 0.0
10:00 201.5 F 0.0
15:00 183.3 F 0.0
20:00 178.4 F 0.0
25:00 180.0 F 0.0
30:00 182.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 26:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:04:28
CYCLE COMPLETE

08/30/2021 02:33 PM
ACCEPTED: X
REJECTED:
Operator: U02