

-----  
Midmark M11 - v1.0.5  
Total Cycles: 8

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2021 08:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.8 F 0.0  
2:00 119.8 F 1.2  
4:00 137.5 F 3.7  
6:00 163.0 F 7.2  
8:00 187.4 F 11.4  
10:00 205.3 F 3.8  
12:00 217.2 F 7.7  
14:00 229.5 F 12.1  
16:00 233.1 F 9.3  
18:00 238.7 F 10.7  
20:00 246.7 F 14.4  
22:00 257.6 F 20.3  
24:00 267.9 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.3  
1:00 272.2 F 30.0  
2:00 271.3 F 29.4  
3:00 272.2 F 29.7  
4:00 271.5 F 29.7

Min 271.3 F 29.2  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 209.8 F 0.0  
10:00 224.0 F 0.0  
15:00 225.3 F 0.0  
20:00 224.0 F 0.0  
25:00 224.0 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 25:10  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

04/22/2021 09:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_