
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001122
8:58 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 3:59
STERILIZING 5:42
136.4 C 223kPa 5:42
136.3 C 224kPa 6:12
136.4 C 225kPa 6:42
136.3 C 224kPa 7:12
136.3 C 223kPa 7:42
136.3 C 224kPa 8:12
136.3 C 224kPa 8:42
136.2 C 222kPa 9:12
136.2 C 222kPa 9:12
VENTING 9:12
AIR DRYING 10:07
CYCLE COMPLETE 25:07

Digital Signature #
D7373E93E143C933
