
Midmark M11 - v1.0.0
Total Cycles: 4623

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2024 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 113.1 F 1.2
4:00 131.6 F 3.7
6:00 161.3 F 7.0
8:00 187.2 F 10.5
10:00 208.2 F 3.7
12:00 213.7 F 7.1
14:00 226.5 F 11.3
16:00 228.2 F 6.9
18:00 239.5 F 12.2
20:00 238.4 F 10.1
22:00 250.9 F 16.0
24:00 261.8 F 22.3
26:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.5
2:00 271.9 F 28.8
3:00 271.6 F 29.0
4:00 272.6 F 29.4

Min 271.3 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 201.7 F 0.0
10:00 202.8 F 0.0
15:00 203.7 F 0.0
20:00 220.4 F 0.0
25:00 232.6 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 26:35
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

06/10/2024 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____