
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007661
9:52 11/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:21
STERILIZING 3:17
136.4 C 220kPa 3:17
136.1 C 223kPa 3:47
136.4 C 220kPa 4:17
136.4 C 223kPa 4:47
136.2 C 224kPa 5:17
136.2 C 223kPa 5:47
136.3 C 220kPa 6:17
136.4 C 222kPa 6:47
136.4 C 222kPa 6:47
VENTING 6:47
AIR DRYING 7:39
CYCLE COMPLETE 23:40

Digital Signature #
5F39913F686CBE75
