

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3302

Name: RITTER 1  
Sterilizer ID: V2178198  
05/30/2023 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
2:00 175.8 F 0.0  
4:00 182.9 F 3.0  
6:00 208.7 F 7.8  
8:00 222.5 F 13.4  
10:00 230.2 F 8.5  
12:00 239.9 F 11.8  
14:00 241.2 F 11.5  
16:00 239.4 F 10.0  
18:00 252.9 F 16.6  
20:00 264.4 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.7  
2:00 271.3 F 28.0  
3:00 272.6 F 28.9  
4:00 272.0 F 28.3  
  
Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 214.5 F 0.0  
10:00 222.4 F 0.0  
15:00 226.2 F 0.0  
20:00 224.3 F 0.0  
25:00 226.6 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 21:60  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:30  
CYCLE COMPLETE

05/30/2023 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_