
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000818
12:25 28/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:29
PRESSURIZING 6:20
STERILIZING 7:51
136.0 C 220kPa 7:51
136.4 C 223kPa 8:21
136.1 C 220kPa 8:51
136.4 C 223kPa 9:21
136.0 C 221kPa 9:51
136.3 C 222kPa 10:21
136.0 C 219kPa 10:51
136.4 C 223kPa 11:21
136.4 C 223kPa 11:21
VENTING 11:21
AIR DRYING 12:23
CYCLE COMPLETE 42:23

Digital Signature #
FA0C86C13BC7BA9A
