
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.7uS / 1.7ppm
CYCLE NUMBER 001873
8:07 19/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:31
STERILIZING 3:38
136.4 C 223kPa 3:38
136.4 C 225kPa 4:08
136.3 C 224kPa 4:38
136.2 C 222kPa 5:08
136.2 C 222kPa 5:38
136.3 C 224kPa 6:08
136.4 C 225kPa 6:38
136.3 C 224kPa 7:08
136.3 C 224kPa 7:08
VENTING 7:09
AIR DRYING 8:00
CYCLE COMPLETE 23:40

Digital Signature #
83806261543CDD4F
