

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1724

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2022 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 184.8 F 1.4  
4:00 195.7 F 6.6  
6:00 213.8 F 12.3  
8:00 229.7 F 8.1  
10:00 241.4 F 11.9  
12:00 243.0 F 12.3  
14:00 242.6 F 11.7  
16:00 256.1 F 18.9  
18:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.6 F 29.3  
2:00 271.7 F 28.6  
3:00 271.7 F 29.1  
4:00 272.7 F 29.3  
  
Min 271.1 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 227.6 F 0.0  
10:00 230.1 F 0.0  
15:00 216.2 F 0.0  
20:00 213.4 F 0.0  
25:00 213.4 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:33  
CYCLE COMPLETE

11/15/2022 11:04 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02