
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 004756
13:25 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 2:56
136.5 C 223kPa 2:56
136.2 C 224kPa 3:26
136.5 C 225kPa 3:56
136.4 C 224kPa 4:26
136.4 C 224kPa 4:56
136.2 C 224kPa 5:26
136.2 C 223kPa 5:56
136.4 C 225kPa 6:26
136.4 C 225kPa 6:26
VENTING 6:26
AIR DRYING 7:19
CYCLE COMPLETE 8:48

Digital Signature #
309CB091D3517C8B
