

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2329

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2022 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.1 F 0.0  
2:00 139.9 F 1.3  
4:00 138.8 F 3.9  
6:00 160.8 F 7.8  
8:00 186.5 F 12.5  
10:00 207.6 F 4.9  
12:00 220.9 F 9.4  
14:00 231.0 F 10.0  
16:00 236.4 F 10.6  
18:00 234.4 F 8.0  
20:00 248.1 F 14.8  
22:00 260.1 F 21.4  
24:00 271.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.8 F 29.5  
2:00 271.8 F 28.8  
3:00 272.9 F 29.6  
4:00 272.2 F 28.9

Min 271.8 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
5:00 197.6 F 0.0  
10:00 197.5 F 0.0  
15:00 202.7 F 0.0  
20:00 202.6 F 0.0  
25:00 194.6 F 0.0  
30:00 187.5 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 24:42  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:03:01  
CYCLE COMPLETE

09/15/2022 09:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_