
Midmark M11 - v1.0.5
Total Cycles: 2732

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 168.1 F 2.9
4:00 189.5 F 7.0
6:00 206.0 F 12.0
8:00 222.8 F 7.3
10:00 235.4 F 12.6
12:00 235.4 F 9.8
14:00 236.2 F 9.0
16:00 247.5 F 14.6
18:00 259.2 F 21.1
20:00 269.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.7
1:00 272.8 F 30.3
2:00 271.7 F 29.9
3:00 272.6 F 30.0
4:00 272.0 F 30.2

Min 271.6 F 29.6
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 216.5 F 0.0
10:00 217.5 F 0.0
15:00 220.7 F 0.0
20:00 223.4 F 0.0
25:00 228.1 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:33
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

08/03/2023 05:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____