

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2061

Name: RITTER2  
Sterilizer ID: V2313228  
02/16/2023 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
2:00 184.5 F 0.1  
4:00 184.1 F 1.9  
6:00 206.3 F 6.4  
8:00 225.2 F 11.3  
10:00 229.2 F 7.3  
12:00 241.4 F 13.0  
14:00 240.6 F 11.2  
16:00 239.3 F 10.2  
18:00 253.0 F 17.1  
20:00 264.6 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.5 F 29.4  
2:00 271.4 F 28.8  
3:00 272.5 F 29.5  
4:00 271.8 F 28.8  
  
Min 271.4 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 225.3 F 0.0  
10:00 227.6 F 0.0  
15:00 229.7 F 0.0  
20:00 225.8 F 0.0  
25:00 226.4 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:05  
CYCLE COMPLETE

02/16/2023 11:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_