
Midmark M11 - v1.0.5
Total Cycles: 1522

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 166.6 F 2.1
4:00 171.6 F 5.9
6:00 194.8 F 11.0
8:00 219.9 F 6.0
10:00 235.0 F 11.4
12:00 231.9 F 8.3
14:00 239.9 F 11.7
16:00 243.6 F 12.7
18:00 255.6 F 19.0
20:00 265.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.9 F 30.3
2:00 271.8 F 29.8
3:00 272.8 F 30.1
4:00 272.3 F 30.4

Min 270.6 F 29.1
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 208.4 F 0.0
10:00 213.1 F 0.0
15:00 215.9 F 0.0
20:00 219.5 F 0.0
25:00 222.1 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

09/22/2022 10:12 AM
ACCEPTED: X
REJECTED:
Operator: U02