
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002046
12:28 16/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:30
136.4 C 225kPa 3:30
136.1 C 222kPa 4:00
136.3 C 224kPa 4:30
136.4 C 226kPa 5:00
136.2 C 222kPa 5:30
136.3 C 224kPa 6:00
136.2 C 224kPa 6:30
136.4 C 225kPa 7:00
136.4 C 225kPa 7:00
VENTING 7:00
AIR DRYING 7:52
CYCLE COMPLETE 25:45

Digital Signature #
F4E736528DDCE4B4
