
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000516
13:57 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:18
136.4 C 232kPa 4:18
136.2 C 232kPa 4:48
136.4 C 233kPa 5:18
136.3 C 233kPa 5:48
136.5 C 232kPa 6:18
136.4 C 232kPa 6:48
136.2 C 231kPa 7:18
136.4 C 233kPa 7:48
136.4 C 233kPa 7:48
VENTING 7:48
CYCLE ABORTED 8:14
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:14

Digital Signature #
BCCD05BFD0D81ADB
