

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1661

Name: RITTER2  
Sterilizer ID: V2313228  
10/26/2022 12:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
2:00 173.3 F 3.6  
4:00 186.3 F 8.0  
6:00 205.8 F 13.1  
8:00 220.8 F 8.3  
10:00 234.8 F 11.6  
12:00 238.6 F 11.6  
14:00 238.2 F 10.3  
16:00 251.6 F 17.0  
18:00 263.2 F 23.8  
20:00 272.4 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.4  
1:00 272.3 F 30.3  
2:00 272.4 F 30.0  
3:00 271.8 F 29.5  
4:00 271.1 F 29.3  
  
Min 271.1 F 29.2  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 227.7 F 0.0  
10:00 233.6 F 0.0  
15:00 226.5 F 0.0  
20:00 226.1 F 0.0  
25:00 227.8 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:56  
CYCLE COMPLETE

10/26/2022 01:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_