
Midmark M11 - v1.0.0
Total Cycles: 938

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2021 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.4 F 0.0
2:00 123.9 F 0.5
4:00 149.2 F 2.2
6:00 171.3 F 4.6
8:00 190.1 F 7.7
10:00 208.8 F 11.8
12:00 216.6 F 5.1
14:00 229.1 F 9.4
16:00 236.3 F 11.2
18:00 238.1 F 10.6
20:00 236.1 F 8.3
22:00 247.7 F 14.3
24:00 258.6 F 20.1
26:00 268.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.7 F 29.3
2:00 272.5 F 28.9
3:00 271.6 F 28.4
4:00 271.7 F 28.8

Min 271.4 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 222.6 F 0.0
10:00 227.8 F 0.0
15:00 220.3 F 0.0
20:00 226.3 F 0.0
25:00 224.2 F 0.0
30:00 221.9 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 27:20
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:05:50
CYCLE COMPLETE

06/24/2021 03:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____