

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3323

Name: RITTER2  
Sterilizer ID: V2313228  
01/25/2024 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.8 F 0.0  
2:00 94.5 F 0.1  
4:00 110.2 F 1.0  
6:00 139.9 F 3.6  
8:00 165.3 F 7.0  
10:00 190.2 F 10.8  
12:00 209.4 F 4.0  
14:00 215.5 F 7.9  
16:00 229.3 F 12.3  
18:00 232.4 F 8.9  
20:00 240.1 F 11.1  
22:00 244.6 F 13.0  
24:00 256.3 F 19.2  
26:00 267.0 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.4 F 29.5  
2:00 271.4 F 29.2  
3:00 272.1 F 29.5  
4:00 271.3 F 29.0

Min 271.3 F 28.9  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 202.8 F 0.0  
10:00 206.9 F 0.0  
15:00 204.3 F 0.0  
20:00 198.4 F 0.0  
25:00 196.9 F 0.0  
30:00 199.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:30  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:05:53  
CYCLE COMPLETE

01/25/2024 03:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_