

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2771

Name: RITTER2  
Sterilizer ID: V2313228  
08/15/2023 02:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
2:00 203.7 F 0.0  
4:00 215.9 F 4.2  
6:00 229.5 F 9.4  
8:00 237.0 F 9.5  
10:00 243.2 F 13.1  
12:00 243.2 F 12.4  
14:00 243.4 F 12.2  
16:00 257.1 F 19.4  
18:00 268.9 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.1 F 28.8  
2:00 272.6 F 29.3  
3:00 271.9 F 29.1  
4:00 272.0 F 28.7  
  
Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 216.1 F 0.0  
5:00 223.3 F 0.0  
10:00 229.8 F 0.0  
15:00 229.2 F 0.0  
20:00 227.7 F 0.0  
25:00 226.8 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 19:01  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

08/15/2023 03:14 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_