
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001644
16:31 30/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:07
STERILIZING 3:08
136.5 C 223kPa 3:08
136.2 C 225kPa 3:39
136.4 C 226kPa 4:09
136.4 C 225kPa 4:39
136.4 C 226kPa 5:09
136.4 C 225kPa 5:39
136.4 C 225kPa 6:09
136.4 C 226kPa 6:39
136.4 C 226kPa 6:39
VENTING 6:39
AIR DRYING 7:30
CYCLE COMPLETE 7:41

Digital Signature #
A2FD50B187759EE2
