
Midmark M11 - v1.0.5
Total Cycles: 3055

Name: RITTER2
Sterilizer ID: V2313228
10/27/2023 08:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.4 F 0.0
2:00 110.5 F 0.1
4:00 114.7 F 0.9
6:00 142.3 F 3.6
8:00 170.3 F 7.6
10:00 194.5 F 11.9
12:00 218.6 F 5.1
14:00 232.0 F 9.7
16:00 233.9 F 9.3
18:00 240.1 F 11.6
20:00 238.0 F 9.8
22:00 250.7 F 16.1
24:00 261.9 F 22.5
26:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.0 F 28.9
2:00 272.3 F 29.1
3:00 272.2 F 29.4
4:00 272.7 F 29.6

Min 271.6 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 237.7 F 0.0
10:00 227.7 F 0.0
15:00 227.6 F 0.0
20:00 228.4 F 0.0
25:00 229.3 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:25
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

10/27/2023 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____