

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5411

Name: RITTER 1  
Sterilizer ID: V2178198  
03/13/2025 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
2:00 168.1 F 2.7  
4:00 176.1 F 6.6  
6:00 194.5 F 10.8  
8:00 205.9 F 4.4  
10:00 219.3 F 8.4  
12:00 231.8 F 13.0  
14:00 231.2 F 9.2  
16:00 240.2 F 10.2  
18:00 244.2 F 12.5  
20:00 255.4 F 18.3  
22:00 266.0 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.1 F 29.4  
2:00 272.5 F 29.1  
3:00 272.3 F 29.5  
4:00 272.6 F 29.3

Min 270.9 F 28.3  
Max 273.3 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.2 F 0.0  
5:00 188.2 F 0.0  
10:00 193.4 F 0.0  
15:00 203.9 F 0.0  
20:00 220.6 F 0.0  
25:00 230.7 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 23:49  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:02:31  
CYCLE COMPLETE

03/13/2025 11:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_