
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001422
13:18 29/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:28
STERILIZING 4:27
136.4 C 224kPa 4:27
136.2 C 223kPa 4:57
136.4 C 225kPa 5:27
136.4 C 225kPa 5:57
136.2 C 222kPa 6:27
136.3 C 223kPa 6:57
136.1 C 222kPa 7:27
136.4 C 224kPa 7:57
136.4 C 224kPa 7:57
VENTING 7:57
AIR DRYING 8:49
CYCLE COMPLETE 8:51

Digital Signature #
A6C898BC111B5A4F
