

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1485

Name: RITTER 1  
Sterilizer ID: V2178198  
12/29/2021 09:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.7 F 0.0  
2:00 127.5 F 0.0  
4:00 159.3 F 1.1  
6:00 187.8 F 3.4  
8:00 205.2 F 7.0  
10:00 220.2 F 11.3  
12:00 222.0 F 4.8  
14:00 233.7 F 9.6  
16:00 237.0 F 9.6  
18:00 241.5 F 11.6  
20:00 238.5 F 9.6  
22:00 251.7 F 16.0  
24:00 262.8 F 22.5  
26:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 29.1  
2:00 272.6 F 28.7  
3:00 271.4 F 28.1  
4:00 271.9 F 28.7

Min 271.2 F 28.1  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
5:00 221.6 F 0.0  
10:00 209.6 F 0.0  
15:00 201.6 F 0.0  
20:00 222.7 F 0.0  
25:00 231.0 F 0.0  
30:00 220.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:04:33  
CYCLE COMPLETE

12/29/2021 10:50 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_