

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2695

Name: RITTER 1  
Sterilizer ID: V2178198  
12/19/2022 05:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 215.8 F 0.0  
2:00 208.7 F 3.7  
4:00 217.8 F 9.0  
6:00 228.2 F 7.2  
8:00 236.1 F 10.7  
10:00 236.3 F 8.2  
12:00 240.8 F 10.2  
14:00 247.4 F 14.1  
16:00 260.2 F 21.2  
18:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.8 F 29.1  
2:00 271.9 F 28.6  
3:00 272.9 F 29.3  
4:00 272.1 F 28.6  
  
Min 271.8 F 28.4  
Max 273.2 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.6 F 0.0  
5:00 234.0 F 0.0  
10:00 230.0 F 0.0  
15:00 228.1 F 0.0  
20:00 230.5 F 0.0  
25:00 230.6 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:52  
CYCLE COMPLETE

12/19/2022 06:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_