

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3037

Name: RITTER2  
Sterilizer ID: V2313228  
10/20/2023 01:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.2 F 0.0  
2:00 124.1 F 1.3  
4:00 153.7 F 4.6  
6:00 181.5 F 9.0  
8:00 198.4 F 8.0  
10:00 227.7 F 7.4  
12:00 237.6 F 12.7  
14:00 236.7 F 10.1  
16:00 235.3 F 8.4  
18:00 248.9 F 15.2  
20:00 261.1 F 22.2  
22:00 272.6 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.5 F 29.4  
2:00 272.0 F 29.8  
3:00 271.8 F 29.1  
4:00 272.5 F 29.9

Min 271.1 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 235.2 F 0.0  
10:00 198.9 F 0.0  
15:00 197.8 F 0.0  
20:00 192.6 F 0.0  
25:00 196.2 F 0.0  
30:00 200.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:24  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:35  
CYCLE COMPLETE

10/20/2023 02:27 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_