
Midmark M11 - v1.0.0
Total Cycles: 2130

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2022 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 171.2 F 1.6
4:00 191.0 F 5.8
6:00 225.6 F 11.4
8:00 223.0 F 6.1
10:00 237.5 F 11.2
12:00 232.3 F 7.7
14:00 244.0 F 12.3
16:00 243.7 F 12.1
18:00 256.4 F 18.6
20:00 267.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.8
2:00 271.3 F 28.3
3:00 272.7 F 28.8
4:00 271.1 F 28.2

Min 271.0 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 235.1 F 0.0
10:00 223.0 F 0.0
15:00 223.5 F 0.0
20:00 222.7 F 0.0
25:00 228.1 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

07/14/2022 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____