

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5300

Name: RITTER 1  
Sterilizer ID: V2178198  
01/30/2025 01:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.7 F 0.0  
2:00 166.1 F 3.4  
4:00 175.4 F 7.7  
6:00 196.7 F 12.1  
8:00 215.3 F 4.9  
10:00 223.0 F 9.6  
12:00 235.4 F 9.6  
14:00 236.5 F 10.4  
16:00 235.9 F 8.2  
18:00 248.1 F 14.4  
20:00 259.6 F 20.6  
22:00 269.8 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.6  
1:00 272.0 F 29.0  
2:00 271.8 F 28.9  
3:00 271.8 F 28.6  
4:00 272.1 F 29.1

Min 271.4 F 28.5  
Max 272.2 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 212.7 F 0.0  
10:00 227.8 F 0.0  
15:00 210.1 F 0.0  
20:00 205.9 F 0.0  
25:00 202.6 F 0.0  
30:00 188.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:01:34  
CYCLE COMPLETE

01/30/2025 02:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_