
Midmark M11 - v1.0.5
Total Cycles: 628

Name: RITTER2
Sterilizer ID: V2313228
12/13/2021 12:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 180.6 F 3.3
4:00 185.9 F 7.7
6:00 204.1 F 12.6
8:00 222.1 F 7.0
10:00 233.9 F 12.1
12:00 235.5 F 9.9
14:00 237.5 F 9.1
16:00 249.0 F 15.5
18:00 261.1 F 22.4
20:00 272.1 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 272.0 F 30.0
2:00 272.3 F 29.6
3:00 271.3 F 29.4
4:00 272.3 F 30.1

Min 271.2 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 223.8 F 0.0
10:00 221.9 F 0.0
15:00 202.0 F 0.0
20:00 195.2 F 0.0
25:00 192.3 F 0.0
30:00 192.1 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

12/13/2021 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____