

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3488

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2023 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.4 F 0.0  
2:00 107.5 F 0.0  
4:00 134.4 F 1.6  
6:00 161.8 F 4.0  
8:00 186.3 F 7.5  
10:00 205.8 F 11.8  
12:00 216.8 F 5.2  
14:00 228.0 F 9.6  
16:00 233.5 F 9.2  
18:00 240.0 F 11.4  
20:00 237.3 F 9.3  
22:00 250.9 F 15.8  
24:00 262.3 F 22.4  
26:00 272.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.7 F 29.1  
2:00 271.6 F 28.3  
3:00 272.3 F 29.2  
4:00 272.2 F 28.9

Min 271.3 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 224.2 F 0.0  
10:00 223.0 F 0.0  
15:00 218.8 F 0.0  
20:00 226.8 F 0.0  
25:00 234.0 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 26:22  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

07/05/2023 01:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_