

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 005128  
16:57 14/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:57  
STERILIZING 2:59  
136.4 C 220kPa 2:59  
136.3 C 221kPa 3:29  
136.4 C 218kPa 3:59  
136.4 C 222kPa 4:29  
136.4 C 225kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 222kPa 5:59  
136.4 C 223kPa 6:29  
136.4 C 223kPa 6:29  
VENTING 6:29  
AIR DRYING 7:17  
CYCLE COMPLETE 10:01

-----

Digital Signature #  
C1603801CEB9911F  
-----