
Midmark M11 - v1.0.0
Total Cycles: 1400

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2021 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 183.4 F 3.0
4:00 195.2 F 6.9
6:00 213.6 F 12.2
8:00 226.6 F 6.7
10:00 236.7 F 11.9
12:00 236.4 F 9.1
14:00 239.2 F 9.8
16:00 247.7 F 14.0
18:00 259.7 F 20.6
20:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.9
2:00 271.6 F 28.3
3:00 271.6 F 28.7
4:00 272.4 F 28.8

Min 271.0 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 224.3 F 0.0
10:00 212.2 F 0.0
15:00 207.9 F 0.0
20:00 210.5 F 0.0
25:00 213.8 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:00
CYCLE COMPLETE

11/29/2021 01:36 PM
ACCEPTED: X
REJECTED:
Operator: U02