
Midmark M11 - v1.0.0
Total Cycles: 2685

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 171.2 F 3.1
4:00 184.8 F 7.5
6:00 205.1 F 12.7
8:00 222.1 F 7.3
10:00 234.2 F 12.7
12:00 236.2 F 9.8
14:00 236.5 F 8.3
16:00 249.4 F 15.0
18:00 261.9 F 22.1
20:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.3 F 29.3
2:00 271.8 F 28.6
3:00 272.0 F 29.1
4:00 272.3 F 28.9

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 218.4 F 0.0
10:00 231.5 F 0.0
15:00 225.9 F 0.0
20:00 227.4 F 0.0
25:00 230.1 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

12/15/2022 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____