
Midmark M11 - v1.0.5
Total Cycles: 1046

Name: RITTER2
Sterilizer ID: V2313228
04/28/2022 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 184.7 F 3.1
4:00 191.2 F 7.6
6:00 211.5 F 13.1
8:00 225.2 F 8.7
10:00 237.3 F 10.7
12:00 241.3 F 12.4
14:00 241.6 F 11.5
16:00 254.8 F 18.3
18:00 266.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.8
2:00 272.0 F 29.1
3:00 271.7 F 29.4
4:00 272.7 F 29.7

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 210.8 F 0.0
10:00 224.3 F 0.0
15:00 230.9 F 0.0
20:00 230.0 F 0.0
25:00 224.7 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

04/28/2022 10:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____