
Midmark M11 - v1.0.0
Total Cycles: 2177

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.6 F 0.0
2:00 183.7 F 3.9
4:00 192.1 F 9.1
6:00 215.0 F 5.9
8:00 228.6 F 9.0
10:00 239.6 F 10.5
12:00 240.8 F 11.7
14:00 239.0 F 10.1
16:00 252.8 F 17.0
18:00 264.5 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.8 F 29.3
4:00 272.1 F 28.6

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 224.1 F 0.0
10:00 225.5 F 0.0
15:00 230.5 F 0.0
20:00 231.3 F 0.0
25:00 228.6 F 0.0
Dry Cycle Aborted

FILLING: 2:05
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 29:34
TOTAL CYCLE: 00:58:48

07/26/2022 11:31 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____