
Midmark M11 - v1.0.0
Total Cycles: 3458

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2023 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 193.7 F 3.5
4:00 198.6 F 8.1
6:00 223.6 F 13.0
8:00 231.0 F 8.8
10:00 240.2 F 10.7
12:00 241.7 F 11.8
14:00 239.7 F 10.5
16:00 253.6 F 17.3
18:00 265.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.5 F 29.2
4:00 271.8 F 28.6

Min 271.2 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 231.4 F 0.0
10:00 233.5 F 0.0
15:00 228.6 F 0.0
20:00 229.3 F 0.0
25:00 229.5 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

06/28/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____