
Midmark M11 - v1.0.5
Total Cycles: 3741

Name: RITTER2
Sterilizer ID: V2313228
06/27/2024 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.7 F 0.0
2:00 95.5 F 0.1
4:00 101.1 F 1.4
6:00 120.7 F 3.9
8:00 149.4 F 7.1
10:00 171.9 F 10.6
12:00 204.9 F 7.0
14:00 212.2 F 7.1
16:00 225.6 F 11.1
18:00 226.5 F 7.2
20:00 238.6 F 12.3
22:00 238.0 F 10.4
24:00 249.7 F 16.1
26:00 260.0 F 22.0
28:00 269.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 30.0
1:00 272.3 F 30.0
2:00 272.0 F 29.9
3:00 271.9 F 29.9
4:00 272.0 F 30.0

Min 271.6 F 29.9
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 210.3 F 0.0
10:00 221.6 F 0.0
15:00 218.3 F 0.0
20:00 225.0 F 0.0
25:00 229.8 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:32
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:08:08
CYCLE COMPLETE

06/27/2024 12:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____