
Midmark M11 - v1.0.0
Total Cycles: 4663

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2024 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 95.7 F 0.0
4:00 84.6 F 0.1
6:00 115.4 F 0.5
8:00 141.8 F 2.4
10:00 165.9 F 5.0
12:00 186.1 F 7.9
14:00 206.4 F 11.7
16:00 218.3 F 4.5
18:00 225.1 F 8.5
20:00 236.3 F 12.5
22:00 233.0 F 8.2
24:00 244.1 F 13.2
26:00 241.4 F 11.0
28:00 253.6 F 17.1
30:00 264.4 F 23.4
32:00 271.2 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 28.6
2:00 271.7 F 28.6
3:00 272.1 F 28.4
4:00 271.9 F 28.8

Min 271.1 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 210.3 F 0.0
10:00 213.4 F 0.0
15:00 211.1 F 0.0
20:00 222.9 F 0.0
25:00 234.9 F 0.0
30:00 226.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:15
CYCLE COMPLETE

06/20/2024 09:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____