

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3326

Name: RITTER2  
Sterilizer ID: V2313228  
01/26/2024 12:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.1 F 0.0  
2:00 128.1 F 0.7  
4:00 151.2 F 3.2  
6:00 178.1 F 6.8  
8:00 192.9 F 11.1  
10:00 221.8 F 4.9  
12:00 223.3 F 9.6  
14:00 233.8 F 9.9  
16:00 240.7 F 11.3  
18:00 237.7 F 9.7  
20:00 251.0 F 16.0  
22:00 262.3 F 22.5  
24:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.6  
2:00 271.2 F 29.5  
3:00 272.0 F 29.6  
4:00 271.2 F 29.3

Min 271.1 F 29.1  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.2 F 0.0  
5:00 224.3 F 0.0  
10:00 214.4 F 0.0  
15:00 193.5 F 0.0  
20:00 179.9 F 0.0  
25:00 177.8 F 0.0  
30:00 176.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:09  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:46  
CYCLE COMPLETE

01/26/2024 01:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_