
Midmark M11 - v1.0.0
Total Cycles: 3708

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2023 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 71.6 F 0.0
4:00 93.1 F 0.0
6:00 123.4 F 1.2
8:00 147.5 F 3.2
10:00 168.7 F 5.8
12:00 185.9 F 8.6
14:00 204.8 F 11.6
16:00 215.3 F 3.6
18:00 221.7 F 6.3
20:00 230.0 F 9.8
22:00 239.6 F 12.3
24:00 237.3 F 9.7
26:00 240.4 F 10.2
28:00 245.2 F 12.6
30:00 255.8 F 18.1
32:00 265.3 F 23.9
34:00 271.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.5
1:00 272.3 F 28.7
2:00 272.5 F 29.0
3:00 272.0 F 29.0
4:00 271.5 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 189.8 F 0.0
10:00 219.2 F 0.0
15:00 229.9 F 0.0
20:00 179.0 F 0.0
25:00 179.2 F 0.0
30:00 178.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 34:30
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:13:04
CYCLE COMPLETE

08/31/2023 09:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____