
Midmark M11 - v1.0.5
Total Cycles: 3490

Name: RITTER2
Sterilizer ID: V2313228
04/01/2024 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 71.3 F 0.0
4:00 87.2 F 0.0
6:00 118.5 F 1.2
8:00 148.8 F 4.6
10:00 172.1 F 8.7
12:00 195.2 F 13.1
14:00 213.2 F 4.8
16:00 224.8 F 8.6
18:00 236.5 F 13.1
20:00 236.9 F 9.9
22:00 236.7 F 8.8
24:00 248.5 F 14.7
26:00 260.8 F 21.6
28:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.8 F 29.5
2:00 271.7 F 28.7
3:00 272.5 F 29.0
4:00 272.5 F 29.4

Min 271.6 F 28.5
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.0 F 0.0
5:00 200.0 F 0.0
10:00 200.7 F 0.0
15:00 203.4 F 0.0
20:00 202.5 F 0.0
25:00 201.2 F 0.0
30:00 199.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:25
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:06:14
CYCLE COMPLETE

04/01/2024 10:39 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____