
Midmark M11 - v1.0.0
Total Cycles: 1393

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2021 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.1 F 0.0
2:00 191.1 F 3.1
4:00 188.5 F 7.4
6:00 208.2 F 13.0
8:00 230.1 F 7.9
10:00 241.4 F 13.2
12:00 239.7 F 10.4
14:00 235.6 F 8.3
16:00 250.8 F 15.4
18:00 262.6 F 22.3
20:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.5 F 28.4
2:00 271.9 F 28.7
3:00 272.8 F 29.0
4:00 271.9 F 28.9

Min 271.4 F 28.4
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 234.5 F 0.0
10:00 211.4 F 0.0
15:00 208.6 F 0.0
20:00 200.3 F 0.0
25:00 194.6 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

11/24/2021 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____