

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2701

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2022 02:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 217.6 F 0.0  
2:00 216.9 F 5.7  
4:00 222.1 F 12.4  
6:00 229.7 F 8.9  
8:00 239.3 F 9.3  
10:00 244.4 F 12.2  
12:00 246.8 F 13.4  
14:00 260.4 F 20.9  
16:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.4 F 28.7  
2:00 272.2 F 28.9  
3:00 272.2 F 28.6  
4:00 272.0 F 28.7  
  
Min 271.5 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 216.2 F 0.0  
5:00 227.4 F 0.0  
10:00 236.0 F 0.0  
15:00 236.2 F 0.0  
20:00 227.1 F 0.0  
25:00 205.4 F 0.0  
Dry Cycle Aborted

FILLING: 2:02  
HEATING: 16:27  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 27:34  
TOTAL CYCLE: 00:53:15

12/20/2022 03:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_